

- One side of the duct is fitted with a transparent acrylic window for visualization.
- An electric heating element is kept in the vertical tube that in turns heats the tube surface.
- The heat is lost from the tube to the surrounding air by natural convection.
- The

temperature of the vertical tube is measure by temperature sensors and displayed by a Digital Temperature Indicator with multi channel switch.

• The

heat input to the heater is measured by a Digital Ammeter and a Digital Voltmeter and is varied by a variac. The tube surface is polished to minimize the radiation losses.

Experimentation:

• To determine average heat transfer coefficient under natural convection

Utilities required:

- Electricity Supply : 1 Phase, 220 V AC, 2 Amp.
- Table for set-up support

Technical Specifications:

- Diameter: 38 mm (approx)
- Length: 500 mm (approx)
- Heater: Nichrome Wire
- Digital Voltmeter: 0-300 Volt
- Digital Ammeter: 0-2 Amp

- Variac: 0-230 V, 2 Amp
- Digital Temp.Indicator: 0-3000C, with multi-channel switch
- Temperature Sensors: RTD PT-100 type
- Instruction Manual: An english instruction manual will be provided along with the Apparatus

Details:

- Powder coated duct of MS to accommodate the assembly with front window of Acrylic
- The whole set-up is well designed and arranged on a powder coated structure

Naugra Export

Website: www.naugraexport.com, Email: sales@naugraexport.com Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India, Phone: +91-0171-2643080, 2601773