Product Name : Planck's Constant Experiment (By Photoelectric Effect) for High School Science Lab for High School Science Kits Lab

Product Code : HS-LAB0002



Description :

Planck's Constant Experiment (By Photoelectric Effect) for Physics Lab- Determination of Planck's Constant and Work Function of Materials by Photoelectric Effect

It was observed as early as 1905 that most metals under influence of radiation, emit electrons. This

phenomenon was termed as photoelectric emission. The detailed study of it has shown.

1. That the emission process depends strongly on frequency of radiation.

2. For each metal there exists a critical frequency such that light of lower frequency is unable to liberate electrons, while light of higher frequency always does.

3. The emission of electron occurs within a very short time interval after arrival of the radiation and member of electrons is strictly proportional to the intensity of this radiation.

The experimental facts given above are among the strongest evidence that the electromagnetic field is quantified and the field consists of quanta of energy E= hn where n is the frequency of the radiation and h is the Planck's constant. These quanta are called photons.

Further it is assumed that electrons are bound inside the metal surface with an energy ef, where f is called work function. It then follows that if the frequency of the light is such that hn > ef

it will be possible to eject photoelectron, while if hn hn = $\frac{1}{2}(mn2 + ef)$ (1)

which is the famous photoelectrons equation formulated by Einstein in 1905.

The energy of emitted photoelectrons can be measured by simple retarding potential techniques as is done in this experiment. When a retarding potential V0 is used to measure kinetic energy of electrons Ee, we have, $Ee = \frac{1}{2}(mv2) = eV0$

or

V0 = (h/e)(n-f)

So when we plot a graph V0 as a function of n, the slope of the straight line yields h and the intercept of extrapolated point n=0 can give work function f.

THE APPARATUS CONSIST OF THE FOLLOWING :

1. Photo Sensitive Device : Vacuum photo tube.

- 2. Light source : Halogen tungsten lamp 12V/35W.
- 3. Colour Filters : 635nm, 570nm, 540nm, 500nm & 460nm.

4. Accelerating Voltage : Regulated Voltage Power Supply,

Output : ± 15 V continuously variable through multi-turn pot

Display : 3¹/₂ digit 7-segment LED

Accuracy : ± 0.2%

5. Current Detecting Unit : Digital Nanoammeter

It is high stability low current measuring instrument

Range : 1000 mA, 100 mA, 10 mA & 1mA with 100 % over ranging facility

Resolution : 1nA at 1 mA range

Display : 31/2 digit 7-segment LED

Accuracy : ±0.2%

6. Power Requirement : $220V \pm 10\%$, 50Hz.

7. Optical Bench : The light source can be moved along it to adjust the distance between light source and phototube scale length is 400 mm. A drawtube is provided to install colour filter, a focus lense is fixed in the back end.

The set is complete in all respect, no additional accessory required.

Naugra Export

Website: www.naugraexport.com, Email: sales@naugraexport.com

Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India, Phone: +91-0171-2643080, 2601773